

CRYOcheck™
Factor X Deficient Plasma

Product Identification

Catalog Number: **FDP10-10**

Lot Number:

Expiry:

| Assay | Value |
|------------------|-------------|
| PT | 113.3 sec |
| APTT | 99.0 sec |
| Fibrinogen | 3.67 g/L |
| Factor II | 107 % |
| Factor V | 98 % |
| Factor VII | 102 % |
| Factor VIII | 101 % |
| Factor IX | 105 % |
| Factor X | < 1.0 % |
| Factor X Antigen | < 0.01 U/mL |
| Factor XI | 92 % |
| Factor XII | 107 % |

Results may vary depending on materials and methods used.

CE

1.800.267.2796

info@precisionbiologic.com

SOLUTION *assured* 