

CRYOcheck™
Factor XII Deficient Plasma

Product Identification

Catalog Number: **FDP12-10; FDP12-15**

Lot Number:

Expiry:

Assay	Value
PT	11.4 sec
APTT	150.5 sec
Fibrinogen	2.90 g/L
Factor II	100 %
Factor V	109 %
Factor VII	121 %
Factor VIII	86 %
Factor IX	109 %
Factor X	110 %
Factor XI	95 %
Factor XII	< 1.0 %
Factor XII Antigen	< 0.01 U/mL

Results may vary depending on materials and methods used.

CE

1.800.267.2796

info@precisionbiologic.com

SOLUTION assured 